

PACKAGING

Ref.: 101-0436	Cont.: 4x240/4x60 mL
Ref.: 101-0571	Cont.: 4x50/2x25 mL

Store at 2-8°C

CLINICAL SIGNIFICANCE

Alkaline phosphatase is an enzyme present in almost all weaves of the organism, being particularly high in bone, liver, placenta, intestine and kidney. Both increases and decreases of plasma ALP are of importance clinically.

Causes of increased plasma ALP: Paget's disease of bone, obstructive liver disease, hepatitis, hepatotoxicity caused by drugs or osteomalacia.

Causes of decreased plasma ALP: Cretinism and vitamin C deficiency^{1,5,6}.

Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

PRINCIPLE OF THE METHOD

Alkaline phosphatase (ALP) catalyses the hydrolysis of p-nitrophenyl phosphate at pH 10.4, liberating p-nitrophenol and phosphate, according to the following reaction:



The rate of p-nitrophenol formation, measured photometrically, is proportional to the catalytic concentration of alkaline phosphatase present in the sample^{1,2}.

REAGENTS

R 1	Diethanolamine (DEA) pH 10.4	1 mmol/L
Buffer	Magnesium chloride	0.5 mmol/L
R 2	p-Nitrophenylphosphate (pNPP)	10 mmol/L
Substrate		

Optional (not included in the kit)

Contro-N	Ref.: 101-0252	4 x 5 mL	Lyophilized human control serum
	Ref.: 101-0083	20 x 5 mL	
Contro-P	Ref.: 101-0253	4 x 5 mL	Lyophilized human control serum
	Ref.: 101-0084	20 x 5 mL	

PREPARATION

Working reagent (WR)

Mix: 4 vol. (R1) Buffer + 1 vol. (R2) Substrate

Stability: 1 month at 2-8°C or 10 days at room temperature (15-25°C).

STORAGE AND STABILITY

All the components of the kit are stable until the expiration date on the label when stored tightly closed at 2-8°C, protected from light and contaminations prevented during their use.

Do not use reagents over the expiration date.

Signs of deterioration:

- Presence of particles and turbidity.
- Blank absorbance (A) at 405 nm ≥ 1.30.

ADDITIONAL EQUIPMENT

- Spectrophotometer or colorimeter measuring at 405 nm.
- Thermostatic bath at 25° C, 30° C o 37° C (± 0.1° C)
- Matched cuvettes 1.0 cm light path.
- General laboratory equipment.

SAMPLES

Serum or heparinized plasma¹. Use unhemolyzed serum, separated from the clot as soon as possible. Stability : 3 days at 2-8° C.

PROCEDURE

Notes: CHRONOLAB SYSTEMS has instruction sheets for several automatic analyzers. Instructions for many of them are available on request.

1. Assay conditions:
Wavelength: 405 nm
Cuvette: 1 cm light path
Constant temperature 25° C / 30° C / 37° C
2. Adjust the instrument to zero with distilled water or air.
3. Pipette into a cuvette:

WR (mL)	1.2
Sample (µL)	20
4. Mix, incubate for 1 minute.
5. Read initial absorbance (A) of the sample, start the stopwatch and read absorbances at 1 min intervals thereafter for 3 min.
6. Calculate the difference between absorbances and the average absorbance differences per minute (ΔA/min).

CALCULATIONS

$$\Delta A/\text{min} \times 3300 = \text{U/L de ALP}$$

Units: One international unit (IU) is the amount of enzyme that transforms 1 mol of substrate per minute, in standard conditions. The concentration is expressed in units per litre of sample (U/L).

Temperature conversion factors

To correct results to other temperatures multiply by:

Assay temperature	Conversion factor to		
	25° C	30° C	37° C
25° C	1.00	1.22	1.64
30° C	0.82	1.00	1.33
37° C	0.61	0.75	1.00

QUALITY CONTROL

Control sera are recommended to monitor the performance of assay procedures.

If control values are found outside the defined range, check the instrument, reagents and technique for problems.

Each laboratory should establish its own Quality Control scheme and corrective actions if controls do not meet the acceptable tolerances.



REFERENCE VALUES¹

	25° C	30° C	37° C
Children (1-14 years)	< 400 U/L	< 480 U/L	< 645 U/L
Adults	60 - 170 U/L	73 - 207 U/L	98 - 279 U/L

Factors affecting ALP activities in a normal population include exercise, periods of repaired growth in children and pregnancy.

These values are for orientation purpose; each laboratory should establish its own reference range.

PERFORMANCE CHARACTERISTICS

Measuring range: From detection limit of 0.6845 U/L to linearity limit of 1233 U/L.

If the results obtained were greater than linearity limit, dilute the sample 1/10 with NaCl (9 g/L) and multiply the result by 10.

Precision:

	Intra-assay (n=20)		Inter-assay (n=20)	
Mean (U/L)	174	443	175	434
SD	0.72	1.56	6.88	11.9
CV (%)	0.41	0.35	3.93	2.75

Sensitivity: 1 U/L = 0.0003 ΔA/min.

Accuracy: Results obtained using CHRONOLAB reagents (y) did not show systematic differences when compared with other commercial reagents (x). The results obtained using 50 samples were the following:

Correlation coefficient (r): 0.99938.

Regression equation: $y = 1.025x - 1.105$.

The results of the performance characteristics depend on the analyzer used.

INTERFERENCES

Fluoride, oxalate, citrate and EDTA inhibit alkaline phosphatase activity and should therefore not be used as anticoagulants. Haemolyses interferes due to the high concentration of alkaline phosphatase in red cells^{1,2}.

A list of drugs and other interfering substances with acid phosphatase determination has been reported by Young et. al^{3,4}.

BIBLIOGRAPHY

1. Wenger C. et al. Alkaline phosphatase. Kaplan A et al. Clin Chem The C.V. Mosby Co. St Louis, Toronto, Princeton 1984; 1094-1098.
2. Rosalki S et al. Clin Chem 1993; 39/4: 648-652.
3. Young DS. Effects of drugs on Clinical Lab. Tests, 4th ed AACC Press, 1995.
4. Young DS. Effects of disease on Clinical Lab. Tests, 4th ed AACC 2001.
5. Burtis A et al. Tietz Textbook of Clinical Chemistry, 3rd ed AACC 1999.
6. Tietz N W et al. Clinical Guide to Laboratory Tests, 3rd ed AACC 1995.